

QUESTIONS AND ANSWERS

01/09/2021

“Medical Equipment Supply”

1. **Publication reference:** 02/2021 - TELE.DOC
2. **Procedure:** Open (International)
3. **Programme:** Interreg IPA Cross Border Cooperation Programme Croatia – Bosnia and Herzegovina- Montenegro 2014 -2020
4. **Contracting Authority:** General Hospital Kotor

No.	Question	Answer
1.	<p>In accordance with purchase of Medical Equipment Supply, publication ref. 02/2021-TELE.DOC and Instruction to tenderers we ask the following questions for Annex II+III: technical specification + technical offer for LOT 3- Operation block Equipment- In the technical specification for III Trauma- Battery operating drill with accessories + charger + container for sterilization for item number 1 is stated: „Battery operating drill / reamer, made of titanium, 250W, 107-110x47-49x200-210mmx55mm...”</p> <p>Given that battery operating drill for all recognized European manufacturers as Aesculap Ag, Germany has three dimensions, is it possible to expand the range of required dimensions and whether it is acceptable to be offered battery operating drill with following dimension 207 x 121 x 55 mm?</p>	<p><i>It is acceptable to offer stated dimensions. Within the published corrigendum no.2 the Contracting Authority has expanded the range of dimensions for battery operating drill / reamer to 200-210mmx107-121x55mm.</i></p>
2.	<p>In the technical specification for V Trauma - set for large fragments for item number 27 is stated: „US Army set of 2 hooks, dimension 26-28 x 15 mm / 22 x 15 mm, 30-32 x 15 MM, 43 x 15 mm, 220mm”.</p> <p>Due to the competitiveness of the procedure, is it possible to expand the</p>	<p><i>It is acceptable to offer stated dimensions. Within the published corrigendum no.2 the Contracting Authority has expanded the range of dimensions for instrument to 26-28 x 15 mm / 22 x 15 mm, 30-39 x 15 MM, 43 x 15 mm.</i></p>



	<p>range of required dimensions and whether it is acceptable to be offered US Army set of 2 hooks, dimension 26x15 mm / 22 x 15 mm, 39 x 15mm/43 x 15 mm, 220mm?</p>	
<p>3.</p>	<p>In the technical specification for IX Abdominal surgery – Inguinal for item number 5 is stated: „Mixer Tissue Grasping Forceps 320mm” Since that tissue grasping forceps according to Mixer for all recognized European manufacturers as Aesculap Ag, Germany do not have this instrument to offer because it does not exist whether it is acceptable to be offered tissue grasping forceps 320mm according to Kelly?</p>	<p><i>It is acceptable to offer Kelly Tissue Grasping Forceps 320mm. Within the published corrigendum no.2 the Contracting Authority allows both models - MIXTER or KELLY.</i></p>